

Session 5A

# Method for Verifying Measurement Uncertainties Against Your Lab's ISO/IEC 17025 Scope of Accreditation

Michael L. Schwartz



# Overview


- Problem: Uncertainty Calculations made by software don't match the SoA
- Introduction to NCSLI 141 Committee
- Introduction to Metrology Taxonomy
- Introduction to SoA & Editor
- How we solved this problem

# Learning Objectives

- Know what a RESTful call is and how it can be used in metrology.
- How standardization solves business problems.
- What the NCSLI 141 Committee is working on.

# About Cal Lab Solutions

- Veteran owned business, ex-military metrologist
- Core business is **software**
  - Metrology consulting
  - Efficiencies through automation
  - Turn-key system
- Focus on finding the right solution for the customer
  - Created Metrology.NET® in 2015
  - World's largest **MET/CAL**® procedure library
  - **PS-Cal** power sensor calibration solution
  - **C#** web-based asset management system
- **Acquired Cal Lab Magazine in 2011**



# Problem:

Often software will generate an uncertainty value that is lower than a calibration lab's Scope of Accreditation uncertainties.



**SureCAL**

**!=**




**≠**

**N7800/TME**



# NCSLI 141 Committee



## Digital Metrology Overview

# Current Technology SoA

## Current Technology



Calculate in Excel



Copy to Word




Create a PDF

# Current Technology SoA

Liquid-In-Glass Thermometers <sup>3</sup>	32 °F  (68 to 662) °F	0.6 °F  0.52 °F	Comparison to digital thermometer in water bath at fixed point (32 °F)  Dry block calibrator
Type J,K,T,E,R,S,C,U,N	32 °F	0.61 °F	Ice bath/precision thermometer CMART 25 calibrator, Fluke 5502A
Type J Type K Type T Type R Type S Type L Type C Type U Type N	(91) °F to (660) °F (91) °F to (660) °F (91) °F to (400) °F (91) °F to (662) °F (91) °F to (662) °F (91) °F to (662) °F (91) °F to (662) °F	0.43 °F 0.43 °F 0.43 °F 0.43 °F 0.43 °F 0.43 °F	Fluke metrology well CMART 25 calibrator Fluke 5502A
Indirect Verification of Rockwell Testers <sup>3</sup>	Rockwell and Portable Rockwell	HRA: (60.5 to 69) HRA (70 to 79) HRA (80 to 92) HRA  HRBW: (0 to 59) HRBW (60 to 79) HRBW (80 to 100) HRBW  HRC: (20 to 35) HRC (35 to 60) HRC (60 to 80) HRC	0.42 HRA 0.41 HRA 0.29 HRA  1.5 HRBW 0.92 HRBW 0.66 HRBW  0.59 HRC 0.51 HRC 0.47 HRC
			Indirect verification per ASTM E18, E110

“Dry well” Search Using Keywords





# Units of Measure are Ambiguous

## Units of Measure aren't enough!

400 fpm

1.7 g

22 °

101 Nm

98.5 %



## Units are really about “Scale”

Scale or Count of a Quantity

## We need Quantity Definitions

To Exchanges data between systems


Convert between Scales



# Metrology Taxonomy

## Category Hierarchy

- 1) Source / Measure
  - 2) Quantity Measured
  - 3) Sub Category
- Sub Category
- Sub Category



# Metrology Taxonomy

## Example for - Source.Voltage.AC.SineWave

The process of sourcing a sinusoidal AC Voltage signal from a device.

This can be used by any device that can generate AC Sinewave Voltage.

### Required Parameters

- Volts – Volts RMS
- Frequency - Hz

### Optional Parameters

- Impedance
- UUT.Range
- UUT.Input – Input Name of the UUT  
Connecting Point

### Measured Value & Uncertainty

- Volts



# Parameter METADATA

**Metadata is data about the data.**

In Metrology.NET, parameters further define the specifics of a test point.

- **Source.Voltage.Sinewave**
  - Volts= 120 V
  - Frequency= 60 Hz
- **Source.Voltage.Sinewave**
  - Volts= 10 V
  - Frequency= 1 MHz
- **Source.Voltage.Sinewave**
  - Volts= 100 kV
  - Frequency= 60 Hz



# Parameter METADATA



**Metadata is data about the data.**

By adding Metadata to an SoA it becomes easier to search

- **Source.Voltage.Sinewave**
  - Volts= 120 V
  - Frequency= 60 Hz
  
- **Source.Voltage.Sinewave**
  - Volts= 10 V
  - Frequency= 1 MHz
  
- **Source.Voltage.Sinewave**
  - Volts= 100 kV
  - Frequency= 60 Hz

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source <sup>1</sup>	Up to 330 mV (0.3 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 000) V	18 nV/mV + 2.4 μV 12 μV/V + 2.7 μV 11 μV/V + 64 μV 18 μV/V + 0.47 mV 21 μV/V + 4.7 mV	Fhike 5520A/SC1100 Multiproduct Calibrator
AC Voltage – Source <sup>1</sup>	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.88 μV/mV + 37 μV 0.17 μV/mV + 37 μV 0.29 μV/mV + 37 μV 1 μV/mV + 36 μV 3.4 μV/mV + 33 μV 8 μV/mV + 61 μV	Fhike 5520A/SC1100 Multiproduct Calibrator
	(33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.65 μV/mV + 51 μV 0.2 μV/mV + 37 μV 0.29 μV/mV + 38 μV 0.7 μV/mV + 38 μV 1.2 μV/mV + 50 μV 2.1 μV/mV + 92 μV	
	(0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.64 mV/V + 0.27 mV 0.21 mV/V + 0.11 mV 0.31 mV/V + 0.11 mV 0.68 mV/V + 0.11 mV 1.1 mV/V + 0.17 mV 2.5 mV/V + 0.78 mV	
	(3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.64 mV/V + 2.7 mV 0.21 mV/V + 1.1 mV 0.41 mV/V + 1.1 mV 0.89 mV/V + 1.1 mV 2.2 mV/V + 1.6 mV	
	(33 to 330) V 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.64 mV/V + 5.4 mV 0.28 mV/V + 8.8 mV 2.5 mV/V + 5 mV 2.5 mV/V + 7.1 mV 4.5 mV/V + 48 mV	
	(330 to 1 000) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.68 mV/V + 18 mV 0.38 mV/V + 19 mV 0.41 mV/V + 19 mV	



# Free SoA Editor

[https://github.com/CalLabSolutions/Metrology.NET\\_Public](https://github.com/CalLabSolutions/Metrology.NET_Public)



Cals-R-Us

## Company Info

- ▲ TestProcess.Measure.Voltage.DC
  - ▲ Measure DCV
    - All
    - ▷ 200mV
    - ▷ 2V
    - ▷ 20V
    - ▷ 200V
    - ▷ 1000V

Measure DCV  
All

Ranges: 5  
Please select a row by clicking on any of the values.

Range	Resolution	Volts	Constants
200mV	8.5	-0.199999999 to 0.199999999	ppm_IV = 4.5 ppm_Range = 0.5
2V	8.5	-1.999999999 to 1.999999999	ppm_IV = 3 ppm_Range = 0.2
20V	8.5	-19.99999999 to 19.99999999	ppm_IV = 3 ppm_Range = 0.2
200V	8.5	-199.9999999 to 199.9999999	ppm_IV = 4.5 ppm_Range = 0.2
1000V	8.5	-1050 to 1050	ppm_IV = 4.5 ppm_Range = 0.5

## Formula

$\text{Volts} * (\text{ppm\_IV} / 1\text{E}+6) + \text{Volts} * (\text{ppm\_Range} / 1\text{E}+6)$   
.15 \*(4.5/1E+6) + .15 \*(0.5/1E+6)

Volts

**Calculate**

7.50E-007

# SoA Editor /Calculator

- **Source.Voltage.Sinewave**
  - Volts= 120 V
  - Frequency= 60 Hz

**8.2E-2 or 0.082V**

**Formula**

**Volts \* Scale + Floor**  
**120 \* 0.00064 + 0.0054**

Volts

**Calculate** **8.22E-002**

**Company Info**

- ▲ Source.Voltage.AC.Sinewave
  - ▲ Fluke 5520A Normal Output
    - All
      - 33 mV
      - 330 mV
      - 3.3 V
      - 33 V
      - 330 V
      - 1000 V
    - ▾ Source.Voltage.DC

Fluke 5520A Normal Output  
All

Ranges: 31  
Please select a row by clicking on any of the values.

**Volts \* Scale + Floor**  
**120 \* 0.00064 + 0.0054**


Volts

**Calculate** **8.22E-002**

Range	Frequency	Volts	Constants
33 V	20e3 to 50e3	3.3 to 33	Floor = 0.0011 Scale = 0.00089
33 V	50e3 to 100e3	3.3 to 33	Floor = 0.0016 Scale = 0.0022
330 V	45 to 1e3	33 to 330	Floor = 0.0054 Scale = 0.00064
330 V	1e3 to 10e3	33 to 330	Floor = 0.0088 Scale = 0.00028

# SoA Calculated Unc Check

## Digital Transformation





# Real World Example



“BestCal” has the Fluke 5520A on their ISO/IEC 17025/2017 Scope of Accreditation.

They bought a new Fluke 5730A with an accredited calibration.

They CAN'T report better than a 5520A.



## Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source <sup>1</sup>	Up to 330 mV (0.3 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 000) V	18 nV/mV + 2.4 μV 12 μV/V + 2.7 μV 11 μV/V + 64 μV 18 μV/V + 0.47 mV 21 μV/V + 4.7 mV	Fluke 5520A/SC1100 Multiproduct Calibrator
	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.88 μV/mV + 37 μV 0.17 μV/mV + 37 μV 0.29 μV/mV + 37 μV 1 μV/mV + 36 μV 3.4 μV/mV + 33 μV 8 μV/mV + 61 μV	
AC Voltage – Source <sup>1</sup>	(33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.65 μV/mV + 51 μV 0.2 μV/mV + 37 μV 0.29 μV/mV + 38 μV 0.7 μV/mV + 38 μV 1.2 μV/mV + 50 μV 2.1 μV/mV + 92 μV	Fluke 5520A/SC1100 Multiproduct Calibrator
	(0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz	0.64 mV/V + 0.27 mV 0.21 mV/V + 0.11 mV 0.31 mV/V + 0.11 mV 0.68 mV/V + 0.11 mV 1.1 mV/V + 0.17 mV 2.5 mV/V + 0.78 mV	
	(3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.64 mV/V + 2.7 mV 0.21 mV/V + 1.1 mV 0.41 mV/V + 1.1 mV 0.89 mV/V + 1.1 mV 2.2 mV/V + 1.6 mV	
	(33 to 330) V 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.64 mV/V + 5.4 mV 0.28 mV/V + 8.8 mV 2.5 mV/V + 5 mV 2.5 mV/V + 7.1 mV 4.5 mV/V + 48 mV	
	(330 to 1 000) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.68 mV/V + 18 mV 0.38 mV/V + 19 mV 0.41 mV/V + 19 mV	

# SoA Calculated Unc Check



Source.Voltage.AC.SineWave

Voltage= 3.5


Frequency= 100e3

The software generated Unc is  $4.8e-4$ .

When approved, the lab's unc will be  $6.2e-4$ .

But the SoA's CMC is  $9.3e-3$ .

So, we must report  $9.3e-3$ .



# Software in Action

## Manual Entry Mode with Fluke 5520A Uncertainty Calculator

Work Order: 20211001-1 Asset Number: OTS-0001 Serial Number: 1234567 Last Test Ran: N/A

Work Order Parameters

Test Package: Agilent 34401... Manufacturer: Agilent Model: 34401A Options Collection Mode: As Found

Test Package Parameters

Agent: CLS\_MICRO\_07 Status: Manual Calibration Run Type: All Tests Assigned To: Admin Admin

Calibration Report

Search:   Show/Hide Sequence

	Group Name	Description	Taxonomy	Status
<input checked="" type="checkbox"/>	AC Volts Verification	1 Year Test Points	TestProcess.Source.Voltage.AC.Sinewave	2 of 12 Tested

Search:   Manual Entry Mode

	Step	Description	Test Type	Nominal	Lower Limit	Upper Limit	Measured	Uncertainty	Status	Summary
<input checked="" type="checkbox"/>	1	100mV Range	Within Limits <>	10.0 mV @1 kHz	9.954	10.046			<input type="checkbox"/> Not Tested	
<input checked="" type="checkbox"/>	2	100mV Range	Within Limits <>	100.0 mV @1kHz	99.900	100.100			<input type="checkbox"/> Not Tested	
<input checked="" type="checkbox"/>	3	100mV Range	Within Limits <>	100.0 mV @50kHz	99.83	100.17			<input type="checkbox"/> Not Tested	
<input checked="" type="checkbox"/>	4	1V Range	Within Limits <>	1.0 V @1kHz	0.99991	1.00090			<input type="checkbox"/> Not Tested	



# Software in Action

**Green** - The custom calculator's uncertainty was equal to or greater than the SoA

**Grey** - The custom calculator's uncertainty was lower than the SoA & REPLACED

Metrology.Net Test Packages Work Orders Uncertainties System Data Admin Cal Lab Solutions

DC Volts Gain Verification 1 Year Test Points TestProcess.Source.Voltage.DC

AC Volts Verification 1 Year Test Points TestProcess.Source.Voltage.AC.Sinewave

Search:  Manual Entry Mode  Check

Step	Description	Test Type	Nominal	Lower Limit	Upper Limit	Measured	Uncertainty	Status	Summary
1	100mV Range	Within Limits <>	10.0 mV @1 kHz	9.954	10.046	10.000	38.70E-6	Passed	TUR 5.2 to 1. Confidence > 99
2	100mV Range	Within Limits <>	100.0 mV @1kHz	99.900	100.100	100.001	57.00E-6	Passed	TUR 3.7 to 1. Confidence > 99
3	100mV Range	Within Limits <>	100.0 mV @50kHz	99.83	100.17	100.00	17.00E-5	Passed	TUR 3.7 to 1. Confidence > 99
4	1V Range	Within Limits <>	1.0 V @1kHz	0.99991	1.00090	1.00004	32.00E-5	Passed	TUR 2.0 to 1. Confidence 84.9
5	1V Range	Within Limits <>	1.0 V @50kHz	0.99830	1.00170	1.00000	12.70E-4	Passed	TUR 4.7 to 1. Confidence > 99
6	10V Range	Within Limits <>	10.0 V @1kHz	9.9910	10.0090	10.0000	9.202E-2	Passed	TUR 0.1 to 1. Confidence 15.5
7	10V Range	Within Limits <>	10.0 V @50kHz	9.9830	10.0170			Not Tested	
8	10V Range	Within Limits <>	10.0 V @10Hz	9.9910	10.0090			Not Tested	
9	100V Range	Within Limits <>	100.0 V @1kHz	99.910	100.090			Not Tested	
10	100V Range	Within Limits <>	100.0 V @50kHz	99.830	100.170			Not Tested	
11	750V Range	Within Limits <>	750.0 V @1kHz	749.325	750.675			Not Tested	
12	750V Range	Within Limits <>	750.0 V @50kHz	748.725	751.275			Not Tested	

# Software in Action

**Green** - The custom calculator's uncertainty was equal to or greater than the SoA

**Grey** - The custom calculator's uncertainty was lower than the SoA & REPLACED

Metrology.Net Test Packages Work Orders Uncertainties System Data Admin Cal Lab Solutions

Search:  Manual Entry Mode  Check

Step	Description	Test Type	Nominal	Lower Limit	Upper Limit	Measured	Uncertainty	Status	Summary
1	100mV Range	Within Limits <>	10.0 mV @1 kHz	9.954	10.046	10.000	38.70E-6	Passed	TUR 5.2 to 1. Confidence > 99
2	100mV Range	Within Limits <>	100.0 mV @1kHz	99.900	100.100	100.001	57.00E-6	Passed	TUR 3.7 to 1. Confidence > 99
3	100mV Range	Within Limits <>	100.0 mV @50kHz	99.83	100.17	100.00	17.00E-5	Passed	TUR 3.7 to 1. Confidence > 99
4	1V Range	Within Limits <>	1.0 V @1kHz	0.99991	1.00090	1.00004	32.00E-5	Passed	TUR 2.0 to 1. Confidence 84.9
5	1V Range	Within Limits <>	1.0 V @50kHz	0.99830	1.00170	1.00000	12.70E-4	Passed	TUR 4.7 to 1. Confidence > 99
6	10V Range	Within Limits <>	10.0 V @1kHz	9.9910	10.0090	10.0000	9.202E-2	Passed	TUR 0.1 to 1. Confidence 15.5
7	10V Range	Within Limits <>	10.0 V @50kHz	9.9830	10.0170			Not Tested	
8	10V Range	Within Limits <>	10.0 V @10Hz	9.9910	10.0090			Not Tested	
9	100V Range	Within Limits <>	100.0 V @1kHz	99.910	100.090			Not Tested	
10	100V Range	Within Limits <>	100.0 V @50kHz	99.830	100.170			Not Tested	
11	750V Range	Within Limits <>	750.0 V @1kHz	749.325	750.675			Not Tested	
12	750V Range	Within Limits <>	750.0 V @50kHz	748.725	751.275			Not Tested	


# Questions? / Comments

**Booth #118**

**6D Panel Discussion 1:00 PM**

**Gatlin E1**

**Metrology's Digital Transformation**



**Michael L. Schwartz**

**Cal Lab Solutions**

**[mschwartz@callabsolutions.com](mailto:mschwartz@callabsolutions.com)**

# Metrology Taxonomy

The NCLI 141 Committee meets weekly to discuss Metrology Taxonomies, M-Layer and other topics related to Digitizing Metrology

Mondays 2:00 pm Mountain Time

Gotomeeting ID 909-871-373

[https://github.com/CalLabSolutions/Metrology.NET\\_Public](https://github.com/CalLabSolutions/Metrology.NET_Public)

## Taxonomy

🔍 All ▾

### TestProcess.Measure.Capacitance.\_

The process of measuring the capacitance of a device.

### TestProcess.Measure.Current.AC.Sinewave

The process measures the AC Current sourced by a device without knowing the shape of the signal. Values can only be tested in RMS because the shape to the signal is unknown.

### TestProcess.Measure.Current.DC

The process measures the Direct Current sourced by a device.


### TestProcess.Measure.Voltage.AC.Sinewave

This process measures the AC RMS (Root-Mean-Square) voltage sourced by the UUT. Values can only be expressed in RMS because the shape of the signal is not known.

### TestProcess.Measure.Voltage.DC

This test process measures the DC (Direct Current) voltage sourced from the UUT.

Metrology Taxonomy Editor



✕

Home

View All

Add

Edit

Delete

Deprecate

Settings

kg m s A mol K



# Changes in Test Equipment Interaction

